

# Unveiling the Spectrum: Versatile Image Processing Techniques in Bone Fracture Detection - A Comprehensive Review

Sanjay S Tippannavar<sup>1</sup>, Yashwanth S D<sup>2</sup>, Gayathri S<sup>3</sup>, Eshwari A Madappa<sup>4</sup>

Electronics and Communication Engineering, JSS Science and Technology University, Mysuru, India

**Email:** <sup>1</sup>sanjayu2345@gmail.com, <sup>2</sup>yashwanth@sjce.ac.in, <sup>3</sup>sgmurthy\_65@sjce.ac.in, <sup>4</sup>eshwariamadappa@gmail.com

## Abstract

India has more than 3,000 accidents every day, according to a global data analysis released by the World Health Organization. Because the bones are rigid, they provide protection for the vital organs of the human body, such as the heart, brain, lungs, and other internal organs. Treating people who have broken bones requires locating the fractures in their bones. Fragility fractures, which result from weakening bones, are linked to a reduction in quality of life and a loss of independence. One of the most popular methods for identifying problems with the human body's organs and bones is to get X-ray Images. Furthermore, more recent methods like magnetic resonance imaging (MRI), microwave imaging (MWI), and computer tomography (CT) are studied. It is the goal of this review paper to educate the research community about the different kinds of image processing techniques available for bone fracture detection and the facilities that hospitals have available for accurate diagnosis. It also serves as a guide for future researchers to understand the procedures needed for the detection of bone fracture.

**Keywords:** Image Processing, Signal Processing, Health, Safety, Artificial Intelligence, Communication, Biomedical, Electronics, Machine Learning

## 1. Introduction

The key organs of the human body, including the heart, brain, lungs, and other internal organs, are protected by the structural pillars of the body—the bones—because of their stiffness. A human body has 206 different types of bones that vary greatly in size, shape, and artifact. In humans, fractures of the bones happen rather often. According to the WHO study, a large number of severe injuries result in death or long-term impairment. It is often quite important for accident situations in many locations where there are few orthopedics accessible. In addition to taking a lot of time, the manual fracture identification method has a high mistake rate. Nowadays, there is a rising tendency in all industries for computer-based methods of defect identification. AI is being used in radiology for a wide range of activities, such as automated illness identification, segmentation, classification, quantification, and many more. Studies reveal that among the many subsets of artificial intelligence (AI), deep learning (DL) is able to identify illnesses from medical images with greater accuracy than doctors. Medical professionals may now provide prompt therapies by evaluating medical images quickly, thanks to developments in computer-aided diagnostic technologies over the last several decades. Bone fractures can be classified based on various factors, including the cause of the fracture, the location of the fracture, the pattern of the break, and the stability of the surrounding structures.

### 1.1. Types of Bone Fractures

1. **Open (compound) Fracture:** The bone breaks through the skin, exposing the fracture site to the external environment. This type of fracture carries an increased risk of infection.
2. **Closed (simple) Fracture:** The skin is intact, but the bone is fractured. The fracture site is not exposed to the environment.
3. **Greenstick Fracture:** Often occurs in children when a bone partially fractures and bends. similar to breaking a green twig.
4. **Transverse Fracture:** The long axis of the bone is parallel to the horizontal fracture line.
5. **Oblique Fracture:** The fracture line is slanted or diagonal across the bone.
6. **Spiral Fracture:** The fracture line encircles the bone and is frequently caused by a twisting force.

7. **Comminuted Fracture:** A bone with at least two fractures is referred to as having a comminuted fracture. Serious injuries such as auto accidents might result in comminuted fractures.
8. **Impacted Fracture:** A break when the ends are forced into one another is known as an impacted fracture, often referred to as a buckle fracture.
9. **Compression Fracture:** A particular kind of fracture called a compression fracture has the potential to shorten your vertebrae by causing them to collapse.
10. **Avulsion Fracture:** happens when a little piece of bone that is linked to a ligament or tendon separates from the main portion of the bone. The most prevalent places for avulsion fractures in young athletes are the hip, elbow, and ankle.
11. **Stress Fracture:** are minute fissures in a bone. They are brought on by repetitive force, frequently from overuse, such as when you run long distances or jump up and down a lot. Normal use of a bone that has been compromised by an illness like osteoporosis can also result in stress fractures.
12. **Pathological Fracture:** bone fracture brought on by a reduction in the bone's mechanical resistance to typical mechanical loads due to structural weakness.
13. **Segmental Fracture:** They occur when a bone is fractured in at least two locations, resulting in a section of the bone that is completely divided by the fractures. You can get these fractures in any long bone in your body.
14. **Intra-articular Fracture:** The fracture affects how the bones articulate because it spreads into the joint surface.
15. **Longitudinal Fracture:** The fracture line is aligned with the bone's longitudinal axis.

Treatment strategy and anticipated healing time may vary depending on the particular kind of fracture. Physicians diagnose and categorize fractures with accuracy using a variety of imaging modalities, including CT, MRI, and X-rays.

## 2. Diagnosis Methods

Imaging investigations, physical examinations, and medical histories are often used to diagnose bone fractures. The following are some standard techniques for identifying bone fractures: These are a few typical techniques for identifying bone fractures depicted in Table 1.

**Table 1.** Diagnosis Methods

| Method            | Principle                                                           | Advantages                                                       | Limitations                                                          | Specificity                                                          |
|-------------------|---------------------------------------------------------------------|------------------------------------------------------------------|----------------------------------------------------------------------|----------------------------------------------------------------------|
| X-rays            | Ionizing radiation penetrates tissues; detects bone density changes | Quick, readily available, cost-effective                         | Limited soft tissue visibility; not ideal for subtle fractures       | High specificity for bone-related abnormalities                      |
| CT Scan           | X-ray technology; provides detailed 3D images of bones and tissues  | Excellent for complex fractures, detailed imaging                | Involves ionizing radiation; higher cost and longer time than X-rays | High specificity for detailed assessment of bony structures          |
| MRI               | Uses magnetic fields and radio waves to produce detailed images     | Excellent soft tissue visualization; no radiation exposure       | Expensive, time-consuming; limited availability in some settings     | High specificity for soft tissue and bone-related abnormalities      |
| Ultrasound        | Uses sound waves to create images of internal structures            | Non-invasive, no radiation exposure; useful for paediatric cases | Operator-dependent; limited use in some fracture types               | Low to moderate specificity; limited by operator expertise           |
| Microwave Imaging | Uses microwave signals to create images of tissues                  | Non-ionizing, potential for real-time imaging                    | Limited penetration; experimental, not widely available              | Limited data; evolving technology                                    |
| Bone Scintigraphy | Radioactive tracer accumulates in areas of increased bone activity  | Sensitive for detecting bone abnormalities                       | Low specificity; exposes patient to radiation                        | Moderate specificity; useful for identifying various bone conditions |

|                   |                                                                     |                                                             |                                                                     |                                                         |
|-------------------|---------------------------------------------------------------------|-------------------------------------------------------------|---------------------------------------------------------------------|---------------------------------------------------------|
| Nuclear Medicine  | Uses small amounts of radioactive materials for diagnostic purposes | Detects bone abnormalities, particularly metabolic activity | Radiation exposure; limited spatial resolution                      | Moderate specificity; useful for certain bone disorders |
| Bone Densitometry | Measures bone mineral density using X-rays                          | Assess bone density; useful for osteoporosis evaluation     | Limited to assessing bone density; not ideal for fracture detection | High specificity for assessing bone mineral density     |

The age of the patient, the suspected nature and location of the fracture, and the accessibility of imaging equipment all influence the diagnostic approach that is used. A thorough knowledge of the fracture may often be attained by combining a number of techniques.

56.2 million individuals worldwide and 10.2 million Americans suffer from osteoporosis, a chronic, progressive bone disease linked to loss of bone density and quality. Physicians must use patient data, especially medical imaging like computed tomography (CT) pictures, to plan therapy and establish diagnoses. This data, together with that of other patients, is very important. Physicians can quickly assess pelvic CT scans and determine the extent of damage with the use of automated fracture diagnosis from segmented bones in traumatic pelvic injuries [1 – 5].

With dedicated extremity coils (knee coil, head coil, or flexible surface coils), MR exams were conducted on a 1.0 T machine. The Salter-Harris classification system was used to categorize the fractures that the MRI indicated. Numerous simulated and real-world MRI bone images have been processed using the SFCM method. The SFCM method performs better at image segmentation when compared to some of the FCM-based algorithms, according to a comparison of these inquiry outcomes [6–8].

This study examines the viability of utilizing microwave imaging (MWI) using a simple and useful setup to identify fractures in superficial bones such as the tibia. In the 8.3–11.1 GHz frequency band, a single Vivaldi antenna is used to capture the dispersed fields by linearly scanning the bone. The Kirchhoff migration technique is used to recreate the image. To eliminate background and skin artifacts, the singular value decomposition (SVD) technique is used. The general concept involves detecting the bone permittivity discontinuity at the fracture

using a scanning microwave monostatic radar device. Nonionizing methods such as microwave imaging (MWI) are becoming more and more popular and show encouraging outcomes [9–10].

Bone X-ray imaging data sets from several publicly accessible research sources, including The Cancer Imaging Archive (TCIA) and the Indian Institute of Engineering Science and Technology, Shibpur, were used for the experiment. The Jupiter laptop running Windows 10 with 8 GB of RAM and an i7 CPU was used for the experiment. Python 3.8 and Anaconda 3 were also used. Bone fracture detection has been feasible with the use of X-ray pictures in developing artificial intelligence approaches, particularly the deep learning method. The deep learning approach proposed in this research study uses X-ray pictures to identify various types of bone fractures and to diagnose bone problems early. In this work, a deep learning-enabled fracture diagnosis method based on X-ray images is developed. To determine which CNN mode is the most accurate, many factors are taken into account [10–15].

The goals of bone fracture detection are: minimizing patient discomfort, making informed decisions, minimizing radiation exposure, accurately diagnosing the condition to guide appropriate interventions, precisely localizing the injury for surgical planning, assessing the severity for individualized care, identifying related injuries, tracking the healing process, preventing complications, customizing rehabilitation programs, and minimizing early identification for prompt treatment.

This review study aims to do the following, principally:

- Analyze the four unique image processing methods that are used to diagnose and identify bone fractures.
- Determine the advantages and disadvantages of the articles, focusing on ways to make them better. This will serve as a useful manual for researchers studying the use of contemporary instruments for the detection of bone fractures. Assess and compile the results of the examined publications.


A range of criteria were used to choose the research papers on bone fracture detection, with a focus on cutting-edge image processing techniques and algorithms. Keywords pertaining to artificial intelligence and machine learning were used in conjunction with terms such as "bone fracture + MRI scan," "bone fracture + CT scan," "bone fracture + MWI," "bone fracture + X-ray," and others while discussing non-invasive screening techniques. Nearly every paper was chosen from journals published by IEEE Xplore, MDPI, and Springer in recent years, spanning from 2021 to 2023, which is testament to the high caliber of the works.

### 3. Methodology

This section covers the use of different software processing algorithms in four non-invasive image processing approaches for the diagnosis of bone fractures. To make the articles simpler to understand, rank, and use in future research, they are organized based on image processing and scanning techniques.

#### 3.1 Bone Fracture Detection using Computer Tomography (CT)


Accurate bone fracture detection with CT scans requires the following: the ability to visualize high-resolution details; the ability to scan quickly to minimize motion artifacts; sophisticated reconstruction algorithms for 3D rendering; appropriate patient positioning and immobilization; the best possible contrast administration when necessary; and competent radiological interpretation. Furthermore, to get accurate diagnostic data for efficient fracture diagnosis and treatment planning, a dependable CT scanner with cutting-edge technology and attention to safety procedures is essential. Physicians can quickly assess pelvic CT scans and determine the extent of injuries with the major assistance of automated fracture identification from segmented bones. This article suggests an automated hierarchical approach for bone fracture identification in pelvic CT scans [1] by combining anatomical information with adaptive windowing, boundary tracing, and wavelet transform. 228 CT images were used for training, 56 for validation, and 408 for testing a method that included two convolutional neural networks (U-Net and ResNet) [2]. In this article, researchers suggest an automated method for identifying cervical spine fractures in CT axial images: a deep convolutional neural network (DCNN) with a bidirectional long-short term memory (BiLSTM) layer. When comparing normal and fractured pictures from CT scans of the cervical spine, there is a significant imbalance. The BiLSTM model has  $N$  axial images and 128 layers, as shown in Figure 1, for the purpose of detecting bone fractures [4].



**Figure 1.** BLSTM Model-based Bone Fracture Detection [1].

### 3.2 Bone Fracture Detection using Magnetic Resonance Imaging (MRI)


A high-field-strength MRI machine with specialized musculoskeletal sequences that provide comprehensive imaging is essential for diagnosing bone fractures using MRI effectively. For motion artifacts to be avoided, patients must be positioned correctly and immobilized. Experts in musculoskeletal imaging must interpret radiological data with skill. A prospective MRI examination was conducted on twenty-four children with acute joint injuries (ten girls and fourteen boys; mean age, 10.7 years; age range, 3–15 years). The Salter-Harris classification system was used to categorize fractures that the MRI showed [6]. The steps needed to detect a bone fracture using MRI are shown in a simple block diagram in Figure 2.



**Figure 2.** Detection of Bone Fracture using MRI [6].

### 3.3 Bone Fracture Detection using Microwave Imaging (MWI)


A nonionizing method that is gaining popularity and showing progressively encouraging outcomes is microwave imaging (MWI). Specialized equipment with the right frequency ranges for bone penetration is needed for the efficient use of microwave imaging in the identification of bone fractures. Accurate data capture depends on an imaging system with well-designed antennas and sensors. The Accuracy of diagnosis is improved by signal processing algorithms that can distinguish between bone structures that are broken and those that are healthy. This work investigates, utilizing a simple and useful setup, the viability of microwave imaging (MWI) for fracture diagnosis in superficial bones such as the tibia. Using a single Vivaldi antenna, the scattered fields are gathered by linearly scanning the bone in the 8.3–11.1 GHz frequency range. A hardware device configuration used to detect bone fractures is shown in Figure 3. The Kirchhoff migration technique is used to recreate the image. To eliminate background and skin artifacts, the singular value decomposition (SVD) technique is used [9–10].



**Figure 3.** Microwave Imaging Hardware Setup for Detection of Bone Fracture [9].

### 3.4 Bone Fracture Detection X-ray

A top-notch X-ray machine that produces detailed pictures, proper patient placement for best visibility, and consistent exposure settings are necessary for accurate bone fracture identification utilizing X-rays. It is essential to have knowledgeable radiologists with experience in musculoskeletal imaging evaluate radiological data. Three separate sources were utilized to obtain the X-ray images. X-ray images from Vidhya Imaging Center, Gwalior, are used to construct the dataset. The biomedical laboratory at the Electrical Engineering Department of MITS, Gwalior, has a few portable digital X-ray machines that are used to capture a few images. An image collection of around 2000 X-ray images has been gathered using the Medpix repository [12]. The techniques for a CAD system used in this research allows for automated fracture diagnosis inside the segmented areas as well as automatic long bone segmentation from X-ray images [15].



**Figure 4.** X-ray based Bone Fracture Detection using CNN [15].

## 4. Analysis of the Articles Reviewed

This section provides an in-depth analysis of the complete outcomes of four different scanning and image processing-based approaches for detecting bone fractures. As an alternative, they are summed based on the dataset, precision, and inaccuracy of the software techniques used to diagnose bone fractures.


### 4.1 Bone Fracture Detection using Computer Tomography (CT)

The dataset was collected from the Virginia Commonwealth University Medical Center. Twelve individuals with severe pelvic injuries have had their data gathered. A total of 45 to 75 images are obtained for every subject. A Registered Active Shape Model (RASM) is used in fracture diagnosis to evaluate the results of prior pelvic bone segmentation [1]. CT scans from 222, 56, and 408 scans were used to train, evaluate, and test an algorithm that included two convolutional neural networks (U-Net and ResNet). In the detection and classification of mandibular fractures on CT, CNNs demonstrated similar reliability and accuracy [2]. Using CT scans of the chest, belly, and pelvis, a deep neural network model has been developed in this research to identify OVFs. This model performed well on many evaluations, in line with the results of radiologists in practice who were using our test set. The best-performing OVF detection approach combined an RNN sequence classifier with a CNN-based neural network to extract features from each CT slice and produce a diagnosis based on the entire CT scan [3–4]. The research using computer tomography to identify bone fractures is summarized in Table 2.

**Table 2.** Parameters for Detection of Bone Fracture using Computer Tomography.

| Cited Work / Parameters | [1]                                             | [2]                                                    | [3]                                                                | [4]                                 |
|-------------------------|-------------------------------------------------|--------------------------------------------------------|--------------------------------------------------------------------|-------------------------------------|
| Dataset                 | Virginia Commonwealth University Medical Center | Peking University School and University of Stomatology | Dartmouth-Hitchcock Medical Center, Lebanon and Montage Healthcare | Cervical Spine CT scans from Kaggle |

|                            |       |                     |                            |                 |
|----------------------------|-------|---------------------|----------------------------|-----------------|
|                            |       |                     | solutions,<br>Philadelphia |                 |
| Accuracy (%)               | 91.98 | 93.87               | 89.2                       | 80.01           |
| Error rate (%)             | 8.02  | 6.13                | 10.8                       | 20              |
| Sensitivity (%)            | 93.33 | 95.95               | 85.2                       | 77.3            |
| Specificity (%)            | 89.26 | 91.40               | 95.8                       | 80.06           |
| Number of Epochs           | NA    | 30                  | 400                        | 50              |
| Batch size                 | NA    | 32                  | 192                        | 16              |
| Optimizer                  | NA    | Adam                | Adam                       | Adam            |
| Learning Rate              | NA    | 10-3                | 10-5                       | 10-3            |
| Hardware employed          | -     | 16-Slice CT scanner | Nvidia Titan Xp GPU Card   | -               |
| Algorithms / Software req. | RASM  | UNet, ResNet        | LSTM, CNN, RNN, ResNet34   | BLSTM, ResNet50 |



**Figure 5.** Comparison of the Hyperparameters in the Articles based on Computer Tomography

In the comparison of the cited works from Figure 5 for bone fracture detection based on computer tomography, several key hyperparameters were considered, including accuracy, specificity, sensitivity, batch size, and the number of epochs. Notably, [2] achieved the highest accuracy (93.87%) and demonstrated strong performance in both sensitivity (95.95%) and

specificity (91.40%). The combination of these metrics indicates a well-balanced model with the ability to effectively identify both positive and negative cases. [3], despite having a lower accuracy (89.2%), stands out with the highest specificity (95.8%), suggesting a lower rate of false positives. On the other hand, [4] exhibits the lowest accuracy (80.01%) and sensitivity (77.3%), which might indicate challenges in identifying positive cases. Regarding training parameters, [3] employed a substantial number of epochs (400) and a large batch size (192), potentially contributing to its higher specificity. However, the choice of the "best" method depends on the specific goals of the application. If a balance between sensitivity and specificity is crucial, [2] might be preferred, while [3] could be chosen for scenarios where minimizing false positives is a top priority. It's essential to consider the trade-offs between these metrics based on the specific requirements of the intended use case.

The compared models exhibit varying limitations in sensitivity, overfitting concerns, low accuracy, and small batch sizes. [4] shows suboptimal sensitivity (77.3%) and low accuracy (80.01%), suggesting challenges in identifying fractures. [3] raises overfitting concerns with 400 epochs. [4] utilizes a small batch size (16), potentially limiting training efficiency. Solutions include fine-tuning for enhanced sensitivity, regularization for overfitting, exploring advanced architectures for accuracy improvement, and experimenting with larger batch sizes for efficiency in [4].

## **4.2 Bone Fracture Detection using Magnetic Resonance Imaging (MRI)**

The Salter-Harris classification system was used to categorize fractures that the MRI indicated. Using specialized extremity coils, MR exams were conducted on a 1.0 T machine [6]. There are 200 non-fractured bone images and 300 fractured images out of the 500 total images. Similar to how 200 normal images are utilized for testing and 130 for training [8], Table 3 displays the studies conducted so far on computer tomography-assisted bone fracture diagnosis.

**Table 3.** Parameters for Detection of Bone Fracture using Magnetic Resonance Imaging.

| Cited Work / Parameters    | [7]            | [8]                                     |
|----------------------------|----------------|-----------------------------------------|
| Dataset                    | Kaggle         | Omni Hospitals, Visakhapatnam           |
| Accuracy (%)               | 78             | 85                                      |
| Error rate (%)             | 22             | 15                                      |
| Sensitivity (%)            | NA             | 87                                      |
| Specificity (%)            | NA             | 86                                      |
| Algorithms / Software req. | SCFM, ANN, DWT | Harris Corner detection, MLP, SURF+BPNN |

### 4.3 Bone Fracture Detection using Microwave Imaging (MWI)

This study examines the viability of utilizing microwave imaging (MWI) using a simple and useful setup to identify fractures in superficial bones such as the tibia. In the 8.3–11.1 GHz frequency band, a single Vivaldi antenna is used to capture the dispersed fields by linearly scanning the bone. Impedance-matching immersion liquids are not necessary since the system may be operated in air. Kirchhoff migration is the method used to rebuild the image [9]. The prior inverse scattering approach, which included just 2D measurement data obtained in a single linear scan, has to be generalized for 3D geometry. Table 4 outlines the requirements for using magnetic resonance imaging (MRI) to detect bone fractures.

**Table 4.** Parameters for Detection of Bone Fracture using Microwave Imaging

| Cited Work / Parameters       | [9]             | [10]            |
|-------------------------------|-----------------|-----------------|
| Accuracy (%)                  | 85              | 87              |
| Error rate (%)                | 15              | 13              |
| Hardware employed             | Vivaldi Antenna | Vivaldi Antenna |
| SCR                           | 3.48dB          | 6.4dB           |
| Permittivity ( $\epsilon_r$ ) | 16              | 24              |

In comparing models [9] and [10], both utilizing Vivaldi Antenna hardware, [10] achieves a slightly higher accuracy (87%) but with a marginally elevated error rate (13%). [9], in contrast, strikes a balance with an accuracy of 85% and a lower error rate of 15%. Notably, [10] demonstrates a superior signal-to-clutter ratio (SCR) of 6.4dB while [9] lags with an SCR of 3.48dB. Additionally, [10] exhibits a higher permittivity ( $\epsilon_r$ ) of 24, surpassing [9]'s 16. Solutions involve optimizing the accuracy-error rate trade-off, enhancing SCR through improved antenna design, and aligning material properties [9].


#### 4.4 Bone Fracture Detection using X-ray

Python 3.8 and Anaconda 3 were used in the experiment, along with a Jupiter notebook on Windows 10 with an i7 CPU and 8 GB of RAM. 10% of the samples were used for testing after the deep CNN was trained using 90% of the samples. For 100 epochs, a batch size of 40 was used using the activation function softmax. After that, the activation function Adam was utilized for 100 epochs with a batch size of 40. [11]. A collection of X-ray images has been gathered from the Medpix database, with over 2000 images total. Subsequently, several attempts have been made to develop a trustworthy identification model with the use of deep learning techniques. This has been accomplished using six CNN models [13]. The findings from studies that used X-rays to identify bone fractures are shown in Table 5.

**Table 5.** Parameters for Detection of Bone Fracture using X-ray

| Cited Work / Parameters | [11]                                   | [12]   | [13]                           | [14]   |
|-------------------------|----------------------------------------|--------|--------------------------------|--------|
| Dataset                 | Cancer Imaging Archive, IEST - Shibpur | Kaggle | Vidhya Imaging Centre, Gwalior | Kaggle |
| Accuracy (%)            | 92.4                                   | 95.4   | 89.90                          | 99.5   |
| Error rate (%)          | 7.6                                    | 4.6    | 10.10*                         | 1*     |
| Sensitivity (%)         | NA                                     | 90     | 87                             | NA     |
| Specificity (%)         | NA                                     | 88     | 89                             | NA     |

|                            |      |                                   |      |                                   |
|----------------------------|------|-----------------------------------|------|-----------------------------------|
| Number of Epochs           | 100  | NA                                | 20   | 10                                |
| Batch size                 | 40   | NA                                | 32   | NA                                |
| Optimizer                  | Adam | NA                                | Adam | NA                                |
| Learning Rate              | NA   | NA                                | 10-3 | 10-3                              |
| Hardware employed          | -    | -                                 | -    | -                                 |
| Algorithms / Software req. | DNN  | CNN,<br>Xception,<br>Inception V3 | CNN  | CNN,<br>CrackNet,<br>Faster R-CNN |



**Figure 6.** Comparison of the Hyperparameters in the Articles based on X-ray.

A range of significant hyperparameters, including batch size, the number of epochs, accuracy, specificity, and sensitivity, were considered when comparing the cited research for the X-ray-based bone fracture diagnosis depicted in Figure 6. The technique with the greatest accuracy (99.5%) and the lowest error rate (1%), [14], suggests a reliable and accurate fracture diagnosis model. Furthermore, [12] has the best-balanced performance, exhibiting the greatest specificity (88%) and sensitivity (90%) rates. Notably, [13] made use of the biggest batch size (32), which could have improved training computational efficiency. [11] demonstrated competitive accuracy for a large number of epochs (100). Selecting the most effective approach is contingent upon the particular needs of the application. When it comes to overall precision and low mistake rates, [14] is a clear winner. But if striking a balance between detail and sensitivity is important, [12] stands out as a good choice. The decision must ultimately be in line with the particular objectives and factors of the X-ray-based bone fracture detection application.

The limitations of the comparison models lie in their varying performance metrics. While [14] achieves exceptional accuracy, its lack of reported sensitivity and specificity limits a comprehensive evaluation. [11] and [13] use relatively higher numbers of epochs, potentially raising concerns about overfitting. A solution involves carefully tuning hyperparameters to prevent overfitting and comprehensive reporting of sensitivity and specificity to enhance model evaluation and comparison.

## 5. Discussions

The benefits and drawbacks of the explored approaches across various technologies employed in bone fracture diagnosis are briefly and critically examined in this part, as seen in Table 6.

**Table 6.** Benefits and Drawbacks of the Studied Methodologies.

| Number | Methodology | Advantages                                                                                                                                                                                                                                                                                         | Disadvantages / Limitations                                                                                                                                                                                                                                                                                                               |
|--------|-------------|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
| 1      | CT Scan     | Given that they provide multiplanar, high-resolution images, CT scans are excellent in detecting bone fractures. Particularly in cases of complicated fractures, its rapidity facilitates prompt diagnostic and treatment choices. Planning surgery might benefit from the thorough visualization. | The use of ionizing radiation during CT scans raises questions about health notwithstanding its advantages. In addition to accessibility issues, the expense is greater. However, there are limitations that should be carefully considered in clinical application, including limited soft tissue contrast and probable metal artifacts. |
| 2      | MRI         | Since MRI scans provide great soft tissue contrast without ionizing radiation,                                                                                                                                                                                                                     | MRI scans, on the other hand, are more costly and take longer. Compared to                                                                                                                                                                                                                                                                |

|   |       |                                                                                                                                                                                                                                             |                                                                                                                                                                                                                                                                                 |
|---|-------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|
|   |       | they are useful for detecting bone fractures. In order to help with precise diagnosis and treatment planning, they provide comprehensive pictures that are helpful for evaluating complicated fractures and related soft tissue injuries.   | other imaging modalities, they may not be as accessible. Some factors to take into account include limited supply in emergency situations and sensitivity to metallic artifacts.                                                                                                |
| 3 | MWI   | Non-ionizing and perhaps portable, microwave imaging may be used to diagnose bone fractures. With the ability to quickly and affordably evaluate fractures, especially in field applications, it provides the benefit of real-time imaging. | The shallow penetration depth of microwave imaging, however, may limit its applicability to deeper anatomical areas or certain types of fractures. To maximize both its specificity and sensitivity for broad clinical use, further investigation and advancement are required. |
| 4 | X-ray | The speed, accessibility, and affordability of X-rays make them a popular tool for detecting bone fractures. They provide rapid and efficient fracture visualization, facilitating timely diagnostic and treatment choices.                 | There is a posing risk since X-rays use ionizing radiation. Comparatively speaking to more sophisticated imaging techniques, they provide less comprehensive information and may not be as sensitive for certain fractures, particularly those affecting soft tissues.          |

The kind of fracture, the clinical setting, and the resources at hand all influence the choice of imaging modalities, each of which has advantages and disadvantages. MRIs and CT scans provide more extensive information, but X-rays are often the first-line imaging method because of their accessibility and rapidity. Though it needs further investigation, microwave imaging offers promise.

**Table 7.** Evaluation and Diagnosis of different Imaging Modalities for Detection of Bone Fracture

| <b>Imaging Modality</b> | <b>Diagnosis and Accuracy for Bone Fracture Detection</b> | <b>Evaluation</b>                                                                                                               |
|-------------------------|-----------------------------------------------------------|---------------------------------------------------------------------------------------------------------------------------------|
| X-ray                   | Rapid, cost-effective, widely accessible.                 | Excellent for initial assessment; limited detail on soft tissues. Concerns with ionizing radiation.                             |
| MRI                     | Excellent soft tissue contrast, no ionizing radiation.    | Superior for complex fractures and soft tissue evaluation; slower and more expensive; limited availability in certain settings. |
| Microwave Imaging       | Potential for real-time, non-ionizing imaging.            | Promising for quick assessment, especially in field applications; limited penetration depth and ongoing research required.      |
| CT Scan                 | High-resolution, multiplanar imaging.                     | Superior for detailed visualization of complex fractures; rapid; concerns with ionizing radiation; higher cost.                 |

## 5.1 Current Trends and Future Perspectives of Bone Fracture Detection

### 5.1.1 Current Trends

1. **Artificial Intelligence (AI) and Machine Learning:** Integration of AI algorithms for automated fracture detection and classification, aiding radiologists in interpreting images and improving diagnostic accuracy.
2. **Advanced Imaging Modalities:** A rise in the thorough evaluation of fractures using modern imaging modalities like MRIs and CT scans, which provide insightful information in complicated situations.

3. **Portable Imaging Solutions:** The advancement of transportable and handheld imaging instruments for point-of-care evaluations, enabling speedier judgments about diagnosis and treatment, particularly in emergency situations.
4. **Telemedicine in Fracture Management:** An increasing number of doctors are using telemedicine platforms to conduct remote consultations and provide advice on fracture care without the requirement for in-person attendance.
5. **3D Printing for Surgical Planning:** Anatomical models are produced using 3D printing technology and used to improve surgical planning accuracy for intricate fracture repair treatments.

#### 5.1.2 Future Perspectives

1. **Microwave Imaging Advancements:** On-going studies to enhance the sensitivity and penetration depth of microwave imaging in order to identify bone fractures.
2. **Smart Implants and Wearable Devices:** Investigation of wearable technology and smart implants that include sensors to follow the healing process, diagnose fractures early, and continuously monitor bone health.
3. **Molecular Imaging Techniques:** Research on molecular imaging methods to evaluate the health of bones at the cellular or molecular level, offering information on bone metabolism and early warning signals of fractures.
4. **Precision Medicine in Fracture Care:** individualized fracture diagnosis and treatment plans based on lifestyle and genetic variables for each patient in order to maximize results.
5. **Regenerative Medicine Approaches:** Developments in tissue engineering and stem cell therapy, which have the potential to improve long-term results by promoting bone regeneration and repair.
6. **Integration of Augmented Reality (AR) and Virtual Reality (VR):** AR and VR technology will be used for training and improved vision during surgery, increasing accuracy in fracture care.
7. **Big Data Analytics in Fracture Research:** Using big data to analyze fracture data thoroughly, improving fracture prediction models and providing a clearer picture of epidemiological patterns.

## 6. Conclusion

As a result of the many imaging techniques, each with its own advantages and disadvantages, the field of bone fracture diagnosis has made great strides. Given its speed and convenience, X-ray imaging is still a vital tool for the preliminary evaluation of fractures, especially when it comes to making an immediate diagnosis in an emergency. High-resolution, detailed pictures are what CT scans are best at producing, enabling accurate diagnosis of complicated fractures and directing surgical procedures. Magnetic resonance imaging (MRI) provides a comprehensive view and is especially helpful in evaluating complex fractures and associated soft tissue injuries because of its superior soft tissue contrast and non-ionizing radiation nature. While further research is required to optimize the penetration depth and sensitivity of microwave imaging for widespread clinical application, this non-ionizing, potentially portable, real-time modality shows promise. Excitement surrounds the future of bone fracture detection: developments in microwave imaging technologies, the incorporation of wearables and smart implants for ongoing surveillance, and the investigation of molecular imaging methods. Artificial intelligence is likely to enable faster interpretation times, better diagnostic precision, and automation of fracture detection processes. Precision fracture treatment strategies that take into account each patient's unique genetics and lifestyle are expected as the field progresses toward customized medicine. Finally, a more sophisticated, effective, and patient-specific approach to bone fracture identification is promised by the combination of various imaging modalities with cutting-edge technology, with the goal of enhancing the overall quality of treatment, lowering radiation exposure, and improving outcomes.

## References

- [1] Yadav, D. P., and Sandeep Rathor. "Bone fracture detection and classification using deep learning approach." In 2020 International Conference on Power Electronics & IoT Applications in Renewable Energy and its Control (PARC), pp. 282-285. IEEE, 2020.
- [2] Meena, Tanushree, and Sudipta Roy. "Bone fracture detection using deep supervised learning from radiological images: A paradigm shift." *Diagnostics* 12, no. 10 (2022): 2420.
- [3] Bagaria, Rinisha, Sulochana Wadhwani, and A. K. Wadhwani. "Bone fracture detection in X-ray images using convolutional neural network." In SCRS Conference

- Proceedings on Intelligent Systems, pp. 459-466. Soft Computing Research Society, 2021.
- [4] Thaiyalnayaki, K., L. Kavyaa, and Joshua Sugumar. "Automated Bone Fracture Detection Using Convolutional Neural Network." In *Journal of Physics: Conference Series*, vol. 2471, no. 1, p. 012003. IOP Publishing, 2023.
- [5] Al-Ayyoub, Mahmoud, Ismail Hmeidi, and Haya Rababah. "Detecting Hand Bone Fractures in X-Ray Images." *J. Multim. Process. Technol.* 4, no. 3 (2013): 155-168.
- [6] Wu, Jie, Pavani Davuluri, Kevin R. Ward, Charles Cockrell, Rosalyn Hobson, and Kayvan Najarian. "Fracture detection in traumatic pelvic CT images." *Journal of Biomedical Imaging* 2012 (2012): 1-1.
- [7] Wang, Xuebing, Zineng Xu, Yanhang Tong, Long Xia, Bimeng Jie, Peng Ding, Hailong Bai, Yi Zhang, and Yang He. "Detection and classification of mandibular fracture on CT scan using deep convolutional neural network." *Clinical Oral Investigations* 26, no. 6 (2022): 4593-4601.
- [8] Tomita, Naofumi, Yvonne Y. Cheung, and Saeed Hassanpour. "Deep neural networks for automatic detection of osteoporotic vertebral fractures on CT scans." *Computers in biology and medicine* 98 (2018): 8-15.
- [9] Salehinejad, Hojjat, Edward Ho, Hui-Ming Lin, Priscila Crivellaro, Oleksandra Samorodova, Monica Tafur Arciniegas, Zamir Merali et al. "Deep sequential learning for cervical spine fracture detection on computed tomography imaging." In *2021 IEEE 18th International Symposium on Biomedical Imaging (ISBI)*, pp. 1911-1914. IEEE, 2021.
- [10] Basha, C. M. A. K., Maruthi Padmaja, and G. N. Balaji. "Computer aided fracture detection system." *Journal of Medical Imaging and Health Informatics* 8, no. 3 (2018): 526-531.
- [11] Gufler, Hubert, Christian Georg Schulze, Sabine Wagner, and Lutz Baumbach. "MRI for occult physeal fracture detection in children and adolescents." *Acta Radiologica* 54, no. 4 (2013): 467-472.
- [12] Sinthura, Siva S., Y. Prathyusha, K. Harini, Y. Pranusha, and B. Poojitha. "Bone Fracture Detection System Using CNN Algorithm." In *2019 International Conference on Intelligent Computing and Control Systems (ICCS)*, pp. 545-549. IEEE, 2019.

- [13] Santos, Kesia C., Carlos A. Fernandes, and Jorge R. Costa. "Feasibility of Bone Fracture Detection Using Microwave Imaging." *IEEE Open Journal of Antennas and Propagation* 3 (2022): 836-847.
- [14] Bekkanti, Ashok, Syed Karimunnisa, Subbarao Gogulamudi, K. T. P. S. Kumar, and CMAK Zeelan Basha. "Enhanced computerized bone fracture detection using harris corner detection." In *2020 International Conference on Smart Electronics and Communication (ICOSEC)*, pp. 572-576. IEEE, 2020.
- [15] Santos, Kesia C., Carlos A. Fernandes, and Jorge R. Costa. "Validation of a Compact Microwave Imaging System for Bone Fracture Detection." *IEEE Access* (2023).